

CERTIFICATE OF COMPLIANCE

Certificate Number 20141024-E363398
Report Reference E363398-A3-UL
Issue Date 2014-OCTOBER-24

Issued to: LABLINE EQUIPMENTS PVT LTD
614/1
GIDC
MAKARPURA
VADODARA GJ 390010 INDIA

This is to certify that
representative samples of

Laboratory Use Electrical Equipment

Centrifuge

Swing 60 / CC :- With Digital Cycle Counter

Swing 60 / WCC :- Without Digital Cycle counter

Swing 60 FT / CC :- With Digital Cycle Counter

Swing 60 FT / WCC :- Without Digital Cycle Counter

Note :- FT is Pre-Set Fix Time models

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety:

UL 61010-1 Safety Requirements For Electrical Equipment
For Measurement, Control, And Laboratory Use - Part 1:

General Requirements

CAN/CSA C22.2 NO. 61010-1 Safety Requirements For
Electrical Equipment For Measurement, Control, And

Laboratory Use -- Part 1: General Requirements

Additional Information:

See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.

B. Mahlenz

Bruce Mahlenz, Assistant Chief Engineer, Global Inspection and Field Services

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at www.ul.com/contactus

